# **Contact-Thermography:** a scientific evaluation

Prepared by: Luis Vitetta<sup>1</sup> Marilyn Johnson<sup>2</sup> Avni Sali<sup>3</sup>

Graduate School of Integrative Medicine Swinburne University

- <sup>1</sup> Director of Research
- <sup>2</sup> Research Assistant
- <sup>3</sup> Head of School



SWINBURNE UNIVERSITY OF TECHNOLOGY

#### Contact-Thermography: a scientific evaluation

Graduate School of Integrative Medicine Swinburne University 9 Frederick Street Hawthorn Vic 3122

Telephone:	+ 61 3 9214 5296
Facsimile:	+ 61 3 9214 8009
Email:	lvitetta@swin.edu.au
Website:	www.swin.edu.au/gsim



© 2002

## Abbreviations

СТ	Contact thermography
DVT	Deep Vein Thrombosis
LCCT	Liquid crystal contact thermography
MVD	Microvessel density
US	Ultrasound

#### Summary

We have examined approximately 200 publications that were associated with thermography from the National Centre for Biotechnology Information and catalogued at the National Library of Medicine (www.ncbi.nlm.nih.gov/). Of these we have reviewed 29 articles for their content, methodology and results.

Although we have not identified any level I or level II evidence associated with contact thermography (CT) nevertheless there is ample evidence to support the use and continued evaluation of this adjuvant screening technique.


Thermography is predominantly a risk marker for breast pathology procedure, which is non-invasive and non-compressive. Perhaps the most pertinent evidence comes from an extensive cohort study of 58,000 patients over a 12 year span that made a significant contribution to the evaluation of patients suspected of having breast cancer. Thermography was then labelled a useful and non toxic tool in the detection of breast cancer. The advent of powerful computers and advances in electronic imaging techniques has further enhanced medical thermography. Today there are systems and detectors with high resolution and temperature sensitivity. Computer algorithms can process raw thermal data and produce high quality profiles of thermal images.

We conclude from the evidence that is currently available that contact thermography may constitute a low cost adjuvant intervention that could be a yearly evaluation as part of a routine physical assessment for women who find mammography a painful and difficult procedure. Therefore as soon as a suspicious (positive) breast thermal examination is performed, the appropriate follow-up diagnostic and clinical testing can be ordered. This would include mammography and other imaging tests, clinical laboratory procedures, nutritional and lifestyle evaluation and training in breast self examination. Thermography is a simple, non-invasive, highly accurate, inexpensive form of diagnostic imaging as well as a 'breast friendly' procedure.

## Contact-Thermography: Literature Review

Thermoregulation is the control of body temperature. The liver produces a lot of heat, which is transported around the body by the blood. Normal body temperature in humans is 37°C. Stability and circadian variation in core body temperature are homeostatic responses that have been well documented for many decades.<sup>1</sup> Research in thermal physiology has illuminated several of the deficits present in the understanding of temperature regulation, and while these discoveries are still evolving, existing information provides valuable clues about physiological responses to heat loss or over-heating that could improve clinical assessment and intervention. Hence, in a healthy individual, body temperature is kept constant in a very small range despite large differences in temperature of the surroundings and also those in physical activity. Strict regulation of body temperature, necessary for optimal progress of enzymatic reactions, is developed in all homoiotermic animals that include humans.

The pathophysiology of heat patterns assumes the diseases of the inner organs and irregularities of their functions are reflected in the heat pattern of the epidermis and can be used diagnostically. The physiology of thermodiagnosis is intimately associated with the brain, the parasympathetic and sympathetic nervous systems (Figure 1). Moreover, the skin circulation and its heat value are largely influenced by processes within the body. The physiology and anatomy of the vascular supply to the skin produces a certain temperature pattern that may be within certain limits that are considered normal. Yet humans show strong deviations from this ideal pattern. Such deviations may have underlying chronic illness correlations. For example in neonatal medicine middle ear infections in neonates often present as a unilateral complaint. That is one side of the head and face produces a pattern of local area temperature increases such as reddening of that side of the face whilst the other side remains normal.<sup>2</sup>



Sagittal Section - Medial View

Figure 1 – Brain and autonomic nervous system

Nerves coming from the spinal cord supply the accompanying body segment with skin and internal organs in their sensitive, vegetative and motor development (Figure 2).




Figure 2 - The segmental innervation of the skin - dermatomes

An electronic search of the medical and scientific literature that included Medline and the Cochrane databases for articles and reviews published between 1966 and 2001 was conducted. The key words that were used included: *thermoregulation, contact* 

thermography, contact thermography and chronic diseases, contact thermography and premature births, contact regulation thermography, contact thermography and breast cancer.

In this report all relevant studies were reviewed on two aspects that included methodology and intervention results. When reviewing research methodologies, the patient selection criteria and procedure, the study design and the variability and reliability of the instrumentation were investigated. The quality of studies was rated according to a four point rating system. This rating system is recommended by the Quality of Care and Health Outcomes Committee and has been adapted from the system developed by the US Preventive ServicesTask Force, which has also been adopted by the National Health and Medical Research Council of Australia. The rating system consists of the following:

- Level I Evidence is obtained from a systematic review of all relevant randomised controlled trials meta analyses.
- Level II Evidence is obtained from at least one properly designed randomised controlled clinical trial.
- Level III Evidence is obtained from well designed controlled trials without randomisation or from well designed cohort or case control analytic design studies, preferably from more than one centre or research group or from multiple time series with or without an intervention.
- Level IV Evidence represents the opinions of respected authorities based on clinical experience, descriptive studies or reports of expert committees.

Recently and following the development and advent of precision temperature measuring instruments known as contact thermometers contact thermography has evolved as a significant and reliable diagnostic method.<sup>3</sup> The scientific published evidence has shown that thermography is a very instructive method of showing the interaction between direct changes in heat radiation of the surface of the skin and its relation to reflex processes.<sup>4</sup>

We have identified from the medical/scientific literature 29 studies comprising either a level III or level IV evidence. These studies show significant methodological detail to afford the authors with a critical appraisal of their validity and reliability.<sup>5-33</sup>

In neonatal medicine, recent studies have shown that measuring lateral skin temperature profile of premature infants in incubators with thermography has enabled skin measurements to be carried out without stress caused by direct contact of probes to the skin. The results verified the greater accuracy and specificity for the use of thermography for measuring skin temperature of preterm infants in incubators.<sup>5</sup> Whereas in a further study of skin temperature measurements in premature infants a comparison of infrared telethermography and electric contact thermometry concluded that in clinical practice, infrared thermography and conductive thermometry can be successfully used for determinations of body surface temperature of premature infants, also under the condition of high temperatures and high humidity within an incubator.<sup>15</sup>

Numerous investigators from different geographical regions that include Denmark, France, Germany, Italy, Japan, United Kingdom and the United State have investigated the effects of contact thermography and more recently liquid crystal contact thermography in chronic diseases such as cancer.

Thermography has been shown to have a useful role as an adjuvant diagnostic tool. The reason for the thermal abnormality associated with some breast cancers is unclear, yet published studies show that an abnormal thermogram can be associated with large tumour size, high grade, and lymph node positivity.<sup>9</sup>

Numerous studies investigating breast abnormalities have shown that contact thermography was effective as a diagnostic tool with high sensitivity for detecting breast abnormalities.<sup>7 11 18 21-23 31 33</sup> Breast abnormalities in breast cancer patients were associated with increases in local regional hyperthermia as related to the tumour as well as a useful adjuvant tool for diagnosing suspected breast neoplasms.

Further contact thermography has also been useful in evaluating chemotherapeutic effectiveness by combining CT as an adjuvant diagnostic procedure.<sup>13</sup> In a preliminary report from Japan the significance of nipple hyperthermia in a contact thermogram aided in the prediction of breast cancer extension to the nipple. The results showed that contact thermography could enhance the detection of breast abnormalities.<sup>34</sup> In a case study of a woman with inflammatory breast cancer from Japan that was reported by Usuki and colleagues showed that CT was useful in evaluating preoperative chemotherapy.<sup>21</sup>

In an Italian study of contact thermography and breast development it was reported that CT was useful in evaluating pubertal breast development and in differentiating between premature thelarche and true precocious puberty.<sup>28</sup> Also in Italy in a study sample consisting of 12,000 patients it was shown that CT proved reliable in those patients aged under 30 in detecting benign pathologies or palpable nodules of the breast.<sup>18</sup>

Geshelin and colleagues have reported that CT is extremely useful as a supplementary diagnostic modality that can significantly distinguish between cancer of the breast and benign tumours.<sup>23</sup> Gautherie and Gros who have evaluated a cohort sample of 58,000 women in the USA over 12 years have reported that CT had made a significant contribution to the evaluation of patients suspected of having breast cancer.<sup>33</sup>

In a study from Japan with various cancer patients it was observed and reported that patients with abnormal CT profiles had cancer recurrences at more than one site.<sup>26</sup>

Contact thermography has also been extensively used in numerous other patient types and conditions that include deep vein thrombosis,<sup>6 25 29 30 32</sup> surgical patients that have examined wound healing,<sup>8</sup> skin damage due to chemical irritants,<sup>27</sup> diabetics with foot ulcers,<sup>12 16</sup> orthopaedic patients and those with temporomandibular dysfunction,<sup>19 20</sup> pregnant women with respiratory infections,<sup>10</sup> children with migraine,<sup>14</sup> facial pain and fifth cranial nerve neuralgia,<sup>24</sup> as well as lung cancer patients<sup>17</sup>

# Published studies

Reference	Level	Source	Patient type	Method and outcome
Frankenberger RT et	IV	Total sample – not	Premature infants	CT
al, 1998⁵		stated		
				CT was useful in measuring skin
		Sample – Germany		temperature of preterm infants in incubators
Kohler A et al, 19986	IV	Total sample – 112	Deep vein thrombosis after	Liquid Crystal Contact
			proximal femur fractures	Thermography (LCCT)
		Sample – USA	F	
				LCCT has proved to be a suitable,
				cheap, non-invasive examination
				with a negative prediction value of 94%
Haga S et al, 1996'	IV	Total sample – 43	Breast cancer	CT
1 laga o ot al, 1000		i otai oampio i o		
		Sample – Japan		68% local regional hyperthermic
				area related to the tumour
Horzic M et al, 1996 <sup>8</sup>	IV	Total sample – 30	Surgical patients	СТ
		Sample – USA		The persistence of a wider zone of
		Sample – USA		increased temperature after day 4
				predicts the possibility of wound
				infection and disturbed healing
Sterns EE et al,	IV	Total sample – 420	Invasive ductal breast	Doppler ultrasound (US)
1996 <sup>9</sup>		Assessed – 181	carcinoma	Microvessel density (MVD)
		Sample – Canada		Contact thermography (CT)
		Sample – Canada		Abnormal CT was associated with
				large tumour size, high grade and
				lymph node positivity
Fisher I et al, 1995 <sup>10</sup>	IV	Total sample – not	Pregnant women with	СТ
		stated	tuberculosis	CT was an effective method for
		Sample – USSR		the detection of respiratory
		Campic COOK		diseases in pregnant women
Abe R et al, 1994 <sup>11</sup>	IV	Total sample – not	Breast cancer	СТ
		stated		
		O		Clinically useful adjuvant
		Sample – Japan		diagnostic method for the early detection of breast cancer in some
				patients
Benbow SJ et al,	IV	Total sample – 30	Diabetic patients with foot ulcers	LCCT
1994 <sup>12</sup>				
		Sample – USA		Patients with high plantar foot
				temperatures are at increased risk
Kurihara T et al,	IV	Total sample – 6	Breast cancer	of neuropathic foot ulceration CT
1993 <sup>13</sup>				
		Sample – Japan		Chemotherapeutic effectiveness
				could be better evaluated by
				combining CT as an adjuvant
Wolstein JR et al,		Total sample – 64	Children with migraine	diagnostic procedure
$1993^{14}$		54 cases		
1000		10 controls		Cases: 36% abnormal CT
		[Age matched]		Controls: 10% abnormal CT
Llongolon Latel	11/	Sample – Canada	Dromoturo infonto	
Hanssler I et al, 1992 <sup>15</sup>	IV	Total sample – 6	Premature infants	СТ
1002		Sample – Germany		CT was concluded as a useful
		sample connury		determinant of body surface
				temperature of premature infants
Chan AW et al,	111	Total sample – 68	Diabetic patients with foot ulcers	СТ
1991 <sup>16</sup>		Cases – 35		Marcalantin
		Controls – 33		Mean foot temperatures were
		[Age, gender matched]		significantly higher in cases than in controls
		matoriouj		
	1	Sample – USA		
		bampic 00A		

Reference	Level	Source	Disease stage and time since diagnosis and/or treatment	Method and outcome
Loviagin EV et al, 1991 <sup>17</sup>	IV	Total sample – 42 Lung cancer – 40 Acute pneumonia – 2	Lung cancers and acute pneumonia	CT Local hyperthermia did not depend on a stage of lung cancer
Sforza M et al, 1991 <sup>18</sup>	IV	Sample – USSR Total sample – 12000	Breast abnormalities	CT CT proved reliable in patients
		Sample – Italy		aged under 30 with benign pathologies or palpable nodules
Specchiulli F et al, 1991 <sup>19</sup>	IV	Total sample – 36 Sample – Italy	Orthopaedic patients	CT CT was an important diagnostic modality in these patients with low
Steed PA, 1991 <sup>20</sup>	IV	Total sample – 109	Temporomandibular dysfunction	cost CT
		Sample – USA		CT proved to be a reliable indicator of pretreatment duration of dysfunction (chronicity pattern) in approximately 78% of the cases
Usuki H et al, 1991 <sup>21</sup>	-	Total sample – 1 [Case report]	Breast cancer	CT CT was found useful for an
Yokoe T et al, 1990 <sup>22</sup>	IV	Sample – Japan Total sample – 162	Breast cancer	evaluation of pre-chemotherapy for inflammatory breast cancer CT
	īv	Sample – Japan	breast cancer	CT sensitivity for detecting breast abnormalities was 81.5% and 83.5% for specificity
Geshelin SA et al, 1989 <sup>23</sup>	IV	Total sample – 162 Benign breast – 87 Malignant breast – 75	Breast abnormalities	CT CT may be considered as a supplementary diagnostic sign distinguishing cancer of the breast
Hardy PA et al, 1989 <sup>24</sup>	IV	Sample – USSR Total sample – not stated	Facial pain and fifth cranial nerve neuralgia	from benign tumours. CT CT was useful in detecting
Hoffmann R et al, 1989 <sup>25</sup>	IV	Sample – UK Total sample – 316 Sample – USA	Deep vein thrombosis	idiopathic trigeminal neuralgia LCCT LCCT is well suited as screening
lkeda T et al, 1989 <sup>26</sup>	IV	Total sample – 97	Various cancer	method for DVTs CT
Among T at at 4000 <sup>27</sup>		Sample – Japan		Patients with abnormal CTs had recurrences at more than one site
Agner T et al, 1988 <sup>27</sup>	IV	Total sample – 16 Sample – Denmark	Skin damage due to chemical irritants	CT CT for assessment of skin damage due to experimental irritants showed positive results
Frejaville E et al, 1988 <sup>28</sup>	111	Total sample – 127 Cases – 22 Controls – 105 Sample – Italy	Breast development study	CT CT was useful in evaluating pubertal breast development and in differentiating between premature thelarche and true precocious puberty
Kjaer L et al, 1988 <sup>29</sup>	IV	Total sample – 56 Sample – Denmark	DVT following major hip surgery	CT CT was not of value as a screening test for DVT following major hip surgery

Reference	Level	Source	Disease stage and time since diagnosis and/or treatment	Method and outcome
Nikulin MA et al, 1987 <sup>30</sup>	IV	Total sample – 430	Vascular disease of the lower extremity	СТ
		Sample – USSR		Useful objective assessment in alterations in blood circulation in the lower extremities resulting from a pathological process
Font Sastre et al, 1984 <sup>31</sup>	IV	Total sample – 2523	Breast screening	СТ
		Sample – France		A higher rate of positive global detections with our present analytic criteria for thermography
Jensen C et al, 1983 <sup>32</sup>	IV	Total sample – 69	DVT	СТ
		Sample – Denmark		CT is useful as a screening method
Gautherie M et al, 1980 <sup>33</sup>	III	Total sample – 58000	Breast cancer	СТ
		[Over 12 years]		CT made a significant contribution to the evaluation of patients
		Sample – USA		suspected of having breast cancer

## References

- 1. Holtzclaw BJ. Circadian rhythmicity and homeostatic stability in thermoregulation. Biol Res Nurs 2001; 2(4):221-35.
- 2. Schloss MD. Otitis media: To treat or not to treat? Can Respir J 1999; 6 Suppl A:51A-3A.
- 3. Rost A. [What value has thermography in diagnosis?]. Krankenpfl J 1982; 20(8):21-3.
- 4. Stuttgen G, Eilers J. Reflex heating of the skin and telethermography. Arch Dermatol Res 1982; 272(3-4):301-10.
- 5. Frankenberger RT, Bussmann O, Nahm W, Konecny E, Gortner L. [Measuring lateral skin temperature profile of premature infants in incubators with thermography]. Biomed Tech (Berl) 1998; 43(6):174-8.
- 6. Kohler A, Hoffmann R, Platz A, Bino M. Diagnostic value of duplex ultrasound and liquid crystal contact thermography in preclinical detection of deep vein thrombosis after proximal femur fractures. Arch Orthop Trauma Surg 1998; 117(1-2):39-42.
- Haga S, Watanabe O, Shimuzu T *et al.* Relation between Locoregional Hyperthermic Area Detected by Contact Thermography and the Maximum Density of Tumor Stain Obtained by IV-DSA in Breast Cancer Patients. Breast Cancer 1996; 3(1):33-7.
- 8. Horzic M, Bunoza D, Maric K. Contact Thermography in a study of primary healing of surgical wounds. Ostomy Wound Manage 1996; 42(1):36-8, 40-2, 44.
- 9. Sterns EE, Zee B, SenGupta S, Saunders FW. Thermography. Its relation to pathologic characteristics, vascularity, proliferation rate, and survival of patients with invasive ductal carcinoma of the breast. Cancer 1996; 77(7):1324-8.
- 10. Fisher Iula, Oborotistova AN, Brio GB. [Factors facilitating the development of tuberculosis in pregnant women and puerperae and the role in their examination of contact film thermal indication]. Akush Ginekol (Mosk) 1995; (6):40-3.
- 11. Abe R, Nakayama K, Takada M. [Early detection and diagnosis of breast cancer]. Gan To Kagaku Ryoho 1994; 21(16):2713-9.
- 12. Benbow SJ, Chan AW, Bowsher DR, Williams G, Macfarlane IA. The prediction of diabetic neuropathic plantar foot ulceration by liquid-crystal contact thermography. Diabetes Care 1994; 17(8):835-9.
- Kurihara T, Higashi Y, Suemasu K, Kanoh T, Tabei T, Inoue K. [Usefulness of contact thermography for the evaluation of chemotherapeutic effectiveness in breast cancer]. Gan To Kagaku Ryoho 1993; 20(7):915-20.
- Wolstein JR, Reed MH, Seshia SS, Kubrakovich P, Linsey B, Samuel A. Contact thermography in the diagnosis of childhood migraine. Can J Neurol Sci 1993; 20(3):222-6.
- 15. Hanssler L, Breukmann H. [Measuring skin temperature in premature infants. Comparison of infrared telethermography and electric contact thermometry]. Klin Padiatr 1992; 204(5):355-8.

- 16. Chan AW, MacFarlane IA, Bowsher DR. Contact thermography of painful diabetic neuropathic foot. Diabetes Care 1991; 14(10):918-22.
- 17. Loviagin EV, Mus VF, Litvinov PD, Iakovleva LA. [Possibilities of contact chromatic thermography in the diagnosis of lung cancer]. Med Radiol (Mosk) 1991; 36(4):11-4.
- Sforza M, Ballerini A, Russo R, Carzaniga PL, Vertemati G. [Contact thermography in breast pathology. A critical review]. Minerva Chir 1991; 46(8):375-7.
- 19. Specchiulli F, Mastrosimone N, Laforgia R, Solarino GB. Acute lesions of the lateral ligaments of the ankle. Clinical and radiographic review. Ital J Orthop Traumatol 1991; 17(2):261-8.
- 20. Steed PA. The utilization of contact liquid crystal thermography in the evaluation of temporomandibular dysfunction. Cranio 1991; 9(2):120-8.
- 21. Usuki H, Murakami M, Komatsubara S *et al.* [A case of inflammatory breast cancer well treated with intraarterial infusion chemotherapy--evaluation of therapy by contact thermography]. Gan To Kagaku Ryoho 1991; 18(7):1191-4.
- Yokoe T, Ishida T, Ogawa T, Iino Y, Kawai T, Izuo M. [Role of cancer thermography for detection of breast cancer]. Gan No Rinsho 1990; 36(8):885-9.
- 23. Geshelin SA, Noskin AL, Kravchenko VA. [Contact thermography in the differential diagnosis of benign tumors and cancer of the breast]. Vrach Delo 1989; (8):103-5.
- 24. Hardy PA, Bowsher DR. Contact thermography in idiopathic trigeminal neuralgia and other facial pains. Br J Neurosurg 1989; 3(3):399-401.
- 25. Hoffmann R, Largiader F, Brutsch HP. [Liquid crystal contact thermography--a new screening procedure in the diagnosis of deep venous thrombosis]. Helv Chir Acta 1989; 56(1-2):45-8.
- 26. Ikeda T, Abe O, Enomoto K, Kikuchi K, Fujiwara K. [Contact thermography as a prognostic indicator of breast cancer]. Gan To Kagaku Ryoho 1989; 16(5):2103-8.
- 27. Agner T, Serup J. Contact thermography for assessment of skin damage due to experimental irritants. Acta Derm Venereol 1988; 68(3):192-5.
- 28. Frejaville E, Pagni G, Cacciari E *et al.* Breast contact thermography for differentiation between premature thelarche and true precocious puberty. Eur J Pediatr 1988; 147(4):389-91.
- 29. Kjaer L, Winter Christensen S, Vestergaard A, Bjerg-Nielsen A, Wille-Jorgensen P. Contact thermography as a screening test for deep venous thrombosis following major hip surgery. Acta Radiol 1988; 29(6):649-52.
- 30. Nikulin MA, Savel'ev IuS. [Diagnostic value of thermography in diseases of the vessels of the lower extremities]. Vestn Khir Im I I Grek 1987; 138(6):43-6.
- 31. Font Sastre V Jr, Julia J, Blanes A, Bonilla Musoles F. Breast cancer screening of the high risk population with clinical examination and thermography. A combination of telethermography and plate thermography. Eur J Gynaecol Oncol 1984; 5(2):105-9.
- 32. Jensen C, Knudsen LL, Hegedus V. The role of contact thermography in the diagnosis of deep venous thrombosis. Eur J Radiol 1983; 3(2):99-102.

- 33. Gautherie M, Gros CM. Breast thermography and cancer risk prediction. Cancer 1980; 45(1):51-6.
- 34. Kurihara T, Suemasu K, Higashi Y. [Significance of nipple hyperthermia in contact thermogram for the prediction of breast cancer extension to the nipple: preliminary report]. Nippon Geka Gakkai Zasshi 1993; 94(8):874.